

Australian Standard<sup>®</sup>

**Surface chemical analysis—X-ray  
photoelectron spectroscopy—  
Repeatability and constancy of intensity  
scale**



This Australian Standard® was prepared by Committee CH-016, Spectroscopy. It was approved on behalf of the Council of Standards Australia on 20 September 2006. This Standard was published on 20 October 2006.

---

The following are represented on Committee CH-016:

- Australian Chamber of Commerce
- Australian Institute of Physics
- CSIRO Energy Technology
- Department of Defence (Australia)
- Department of Natural Resources and Mines Qld
- National Association of Testing Authorities Australia
- Queensland Health Scientific Services

Additional Interests:

- Curtin University of Technology
  - La Trobe University
  - University of Newcastle
  - The University of New South Wales
  - The University of Queensland
  - University of South Australia
- 

This Standard was issued in draft form for comment as DR 06412.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through public comment period.

---

### **Keeping Standards up-to-date**

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting **[www.standards.org.au](http://www.standards.org.au)**

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at **[mail@standards.org.au](mailto:mail@standards.org.au)**, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

---

Australian Standard<sup>®</sup>

**Surface chemical analysis—X-ray  
photoelectron spectroscopy—  
Repeatability and constancy of intensity  
scale**

First published as AS ISO 24237—2006.

**COPYRIGHT**

© Standards Australia

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 0 7337 7781 3

## PREFACE

This Standard was prepared by the Standards Australia Committee CH-016, Spectroscopy. This Standard is identical with, and has been reproduced from, ISO 24237:2005, *Surface chemical analysis—X-ray photoelectron spectroscopy—Repeatability and constancy of intensity scale*.

The objective of this Standard is to ensure that the requirements for evaluating repeatability and constancy of the intensity scale of X-ray photoelectron spectrometers for general analytical purposes are achieved.

As this Standard is reproduced from an international standard, the following applies:

- (a) Its number appears on the cover and title page while the International Standard number appears only on the cover.
- (b) In the source text 'this International Standard' should read 'this Australian Standard'.
- (c) A full point substitutes for a comma when referring to a decimal marker.

The term 'informative' has been used in this Standard to define the application of the annex to which it applies. An informative annex is only for information and guidance.

## CONTENTS

|                     | <i>Page</i>                                                                                                                                                                                                         |
|---------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>1</b>            | <b>Scope ..... 1</b>                                                                                                                                                                                                |
| <b>2</b>            | <b>Symbols and abbreviations ..... 1</b>                                                                                                                                                                            |
| <b>3</b>            | <b>Outline of method ..... 2</b>                                                                                                                                                                                    |
| <b>4</b>            | <b>Method for evaluating the repeatability and constancy of the intensity scale..... 2</b>                                                                                                                          |
| <b>4.1</b>          | <b>Obtaining the reference sample ..... 2</b>                                                                                                                                                                       |
| <b>4.2</b>          | <b>Mounting the sample ..... 3</b>                                                                                                                                                                                  |
| <b>4.3</b>          | <b>Cleaning the sample ..... 3</b>                                                                                                                                                                                  |
| <b>4.4</b>          | <b>Choosing the spectrometer settings for which intensity stability is to be determined ..... 4</b>                                                                                                                 |
| <b>4.5</b>          | <b>Operating the instrument ..... 5</b>                                                                                                                                                                             |
| <b>4.6</b>          | <b>Options for initial or subsequent evaluation measurements ..... 5</b>                                                                                                                                            |
| <b>4.7</b>          | <b>Measurements for the intensity and repeatability ..... 5</b>                                                                                                                                                     |
| <b>4.8</b>          | <b>Calculating the peak area intensities, intensity ratios and uncertainties ..... 6</b>                                                                                                                            |
| <b>4.9</b>          | <b>Procedure for the regular evaluation of the constancy of the intensity scale ..... 7</b>                                                                                                                         |
| <b>4.10</b>         | <b>Next evaluation ..... 7</b>                                                                                                                                                                                      |
| <b>Annex A</b>      | <b>(informative) Example of calculations and measurements of the intensity repeatability<br/>for a commercial X-ray photoelectron spectrometer using unmonochromated<br/>Mg K<math>\alpha</math> X-rays ..... 9</b> |
| <b>Bibliography</b> | <b>..... 12</b>                                                                                                                                                                                                     |

## INTRODUCTION

X-ray photoelectron spectroscopy (XPS) is used extensively for the surface analysis of materials. Elements in the sample (with the exception of hydrogen and helium) are identified from comparisons of the measured binding energies of their core levels with tabulations of those energies for the different elements. Information on the quantities of such elements can be derived from the measured photoelectron intensities. Calculation of the quantities present may then be made using formulae and relative sensitivity factors provided by the spectrometer manufacturer. It is important that the sensitivity factors are appropriate for the instrument and this will generally be the case directly after installation of the equipment or calibration of the instrument intensity/energy response function by an appropriate organization. There are two important instrumental contributions to the uncertainty of XPS intensity measurements that are addressed in this International Standard: (i) the repeatability of intensity measurements and (ii) the drift of the intensities with time.

Repeatability is important for analysing the trends and differences between samples that are similar. The instrumental issues that limit the measurement repeatability include the stability of the X-ray source, the settings of the detector, the sensitivity of the instrument to the sample placement, the data acquisition parameters and the data-processing procedure. The drift of the instrument intensity scale will limit the overall accuracy of any quantitative interpretation and arises from such effects as the ageing of components of the structure of the spectrometer, of its electronic supplies and of the detector. In XPS instruments, it has been found that, in service, the instrument intensity/energy response function may change as the instrument ages.

This International Standard describes a simple method for determining the repeatability and constancy of the intensity scale of the instrument so that remedial action, such as improving the operating procedure, resetting of the instrument parameters or recalibration of the intensity/energy response function, may be made. This method should, therefore, be conducted at regular intervals and is most useful if the data include a period in which the instrument has been checked to be working correctly by the manufacturer or other appropriate body. This method uses a sample of pure copper (Cu) and is applicable to X-ray photoelectron spectrometers with unmonochromated aluminium (Al) or magnesium (Mg) X-rays or monochromated Al X-rays.

This method does not address all of the possible defects of instruments since the required tests would be very time-consuming and need both specialist knowledge and equipment. This method is, however, designed to address the basic common problem of repeatability and of drift of the intensity scales of XPS instruments. This method may be conducted at the same time as the spectrometer energy calibration using ISO 15472 [1].

## AUSTRALIAN STANDARD

# Surface chemical analysis — X-ray photoelectron spectroscopy — Repeatability and constancy of intensity scale

## 1 Scope

This International Standard specifies a method for evaluating the repeatability and constancy of the intensity scale of X-ray photoelectron spectrometers, for general analytical purposes, using unmonochromated Al or Mg X-rays or monochromated Al X-rays. It is only applicable to instruments that incorporate an ion gun for sputter cleaning. It is not intended to be a calibration of the intensity/energy response function. That calibration may be made by the instrument manufacturer or other organization. The present procedure provides data to evaluate and confirm the accuracy with which the intensity/energy response function remains constant with instrument usage. Guidance is given on some of the instrument settings that may affect this constancy.

## 2 Symbols and abbreviations

|               |                                                                                                                                                                               |
|---------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| $A_2$         | average peak area for the Cu 2p <sub>3/2</sub> peak after removing the Shirley background                                                                                     |
| $A_{2j}$      | a value contributing to $A_2$ for the $j$ th measurement in a set of measurements                                                                                             |
| $A_3$         | average peak area for the Cu 3p peak after removing the Shirley background                                                                                                    |
| $A_{3j}$      | a value contributing to $A_3$ for the $j$ th measurement in a set of measurements                                                                                             |
| $i$           | identifier for one of the three parameters $P_i$                                                                                                                              |
| $j$           | index for one of the individual measures of the parameter $P_{ij}$                                                                                                            |
| $P_i$         | parameter representing the mean value of any of $A_2$ , $A_3$ and $A_3/A_2$                                                                                                   |
| $P_{ij}$      | the $j$ th measure of parameter with average value $P_i$                                                                                                                      |
| $U_{95}(P_i)$ | uncertainty in the mean value of $P_i$ , at 95 % confidence level                                                                                                             |
| XPS           | X-ray photoelectron spectroscopy                                                                                                                                              |
| $\delta$      | value of the tolerance limit for $A_3/A_2$ for compliance at 95 % confidence level (set by the analyst)                                                                       |
| $\Delta$      | energy offset for the instrumental binding energy scale, equal to the measured Cu 2p <sub>3/2</sub> binding energy value for the maximum intensity at the peak minus 932,7 eV |
| $\sigma(P_i)$ | repeatability standard deviation for the parameter $P_i$                                                                                                                      |