



BSI Standards Publication

Tissue-engineered medical products — Evaluation of anisotropic structure of articular cartilage using DT (Diffusion Tensor)-MR Imaging

National foreword

This Published Document is the UK implementation of ISO/TR 16379:2014.

The UK participation in its preparation was entrusted to Technical Committee RGM/1, Regenerative medicine.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014.
Published by BSI Standards Limited 2014

ISBN 978 0 580 82564 4
ICS 11.040.40

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 May 2014.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

**Tissue-engineered medical
products — Evaluation of anisotropic
structure of articular cartilage using
DT (Diffusion Tensor)-MR Imaging**

*Produits médicaux à base de tissus — Évaluation de la structure
anisotrope du cartilage articulaire en utilisant l'imagerie en tenseur
de diffusion (IRM-TD)*





COPYRIGHT PROTECTED DOCUMENT

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Terms and definitions	1
3 Principle	3
4 Diffusion tensor magnetic resonance imaging (DT-MRI) data observation in articular cartilage	3
4.1 DT-MRI measurement process	3
4.2 Notes on setting of DT-MRI imaging parameters for articular cartilage	4
4.3 Measurement indices for structural evaluation of articular cartilage by DT-MRI	5
Annex A (informative) Measurement Results	10
Bibliography	23

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 150, *Implants for Surgery*, Subcommittee SC 7, *Tissue-engineered Medical Products*.

Introduction

Structural evaluation of articular cartilage with conventional diagnostic technologies is challenging, and Nihon University has developed technologies (see Reference [1]) and collected relevant data for *in vivo* evaluation of articular cartilage structure by means of diffusion tensor magnetic resonance imaging (DT-MRI) using 1,5 Tesla or 3 Tesla MRI equipment employed for treatment in hospital settings. These data are released in this Technical Report prepared for reference in treatment settings.

This work is part of “Development of Cartilage Observation and Evaluation Technologies for Regenerative Medicine Processes”, an activity managed by the University under “Development of Evaluation Technology for Early Introduction of Regenerative Medicine”, a project contracted by the New Energy and Industrial Technology Development Organization (NEDO) to the National Institute of Advanced Industrial Science and Technology (AIST) and its Technology Research Association of Medical Welfare Apparatus.

Tissue-engineered medical products — Evaluation of anisotropic structure of articular cartilage using DT (Diffusion Tensor)-MR Imaging

1 Scope

This Technical Report has been prepared for evaluation of therapeutic courses for articular cartilage disease and summarizes results from structural evaluation of knee joint cartilage by diffusion tensor imaging, an MRI applied technology allowing non-invasive observation of soft tissue morphology *in vivo*.

This Technical Report is intended for use in areas such as regenerative medicine for knee joint cartilage disease.

After *in vivo* transplant of cartilage cells or tissue as a regenerative treatment, longitudinal diagnosis is needed to assess regeneration as articular cartilage, but arthroscopes used primarily for this purpose are invasive and also do not allow evaluation of structure by simple observation of surfacial characteristics. Radiography and CT do not visualize articular cartilage and also entail the problem of exposure. Collagen fibres, the primary component of articular cartilage, have a surfacial layer parallel to the articular surface to serve a lubricating function for the articular surface, a middle layer with a randomized structure to distribute loads, and deep layers oriented vertically to support loads. The anisotropy of this three-layer structure is a characteristic feature of hyaline cartilage structures and a mechanism demonstrating a lubricating function for articular cartilage. We can then ask whether articular cartilage can be assessed by evaluating the anisotropy of collagen.

MRI techniques allow non-invasive visualization of soft tissue form and function *in vivo*, and DT-MRI conveys the direction of water molecule motion. In fibrous tissues, the direction of water molecule motion is restricted to the direction of fibre orientation; consequently, the direction of water molecule motion matches that of fibre orientation. The use of DT-MRI therefore does allow evaluation of collagen fibre orientation and anisotropy in articular cartilage.

DT-MRI is thus used to observe articular cartilage anisotropy data for use as standardized data in longitudinal diagnosis following transplant of articular cartilage as a regenerative treatment.

2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

2.1

diffusion tensor

DT

tensor expressing the orientation and magnitude of diffused proton signals

2.2

sequence

protocol for performance of MRI

2.3

spin-echo echo-planar imaging

SE-EPI

method of high-speed imaging in which gradient fields are flipped continuously at high speed to produce echoes continuously by means of a spin-echo pulse sequence