



BSI Standards Publication

Soil quality — Detection of water soluble chromium(VI) using a ready-to-use test-kit method

National foreword

This Published Document is the UK implementation of ISO/TR 18105:2014.

The UK participation in its preparation was entrusted to Technical Committee EH/4, Soil quality.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014. Published by BSI Standards Limited 2014

ISBN 978 0 580 79525 1

ICS 13.080.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 September 2014.

Amendments issued since publication

Date	Text affected
------	---------------

**Soil quality — Detection of water
soluble chromium(VI) using a ready-
to-use test-kit method**

*Qualité du sol — Détection du chrome(VI) soluble dans l'eau en
utilisant un kit d'essai prêt à l'emploi*





COPYRIGHT PROTECTED DOCUMENT

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 References	1
3 Terms and definitions	1
4 Principle	1
5 Reagents	2
6 Interferences	2
7 Apparatus	2
8 Procedure	2
8.1 Calibration	2
8.2 Sample detection	2
8.3 Measurement	2
9 Quality assurance (QA) and quality control (QC)	3
10 Test report	3
Annex A (informative) Extraction of Cr(VI) from soil CRMs	4
Annex B (informative) Comparison of Cr(VI) results obtained with commercially available test-kits	7
Annex C (informative) Simulation of mapping distributions of Cr(VI) contaminating potential land	8
Bibliography	14

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information.

The committee responsible for this document is ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical methods and soil characteristics*.

Introduction

A test-kit method uses reagents stored in plastic or glass vials prepared for colorimetric detection of specific compounds soluble with water. There is a lot of varieties of test-kits developed mainly for the compounds in water samples based on well-known reactions with commonly-used and chemically stable reagents indicating constant magnitudes of intensities of the signals from target compounds. The common reactions employed for the test-kit usually give colour, by which the user can recognize the concentration level of the compound without absorption spectrophotometers even in the situation when such instruments are not available. Thus, the test-kit determination of the compound in water samples can be used to screen specific compounds in soil samples, if they are easily dissolved into water.

The reaction applied depends on the target compounds. For chromium(VI), a reaction is used, where 1,5-diphenylcarbazide is used, its isomer or derivatives. When other chemicals or derivatives are used, the performance is intended to be similar to that obtained with the original reagents. The test-kit consisting of the reagents has to be identified by confirming the spectrum including intensities at a specific wavelength. The spectrum shape should be similar to each other between those from test-kits and bulk reagents commonly used for absorptiometry in laboratories. The intensities and varieties of interferences to the test-kit method should be within those taking place to absorptiometry test methods with bulk reagents. The quality or purity of the reagents packed in test-kits should keep constant with a deviation around that for the bulk reagents for absorptiometry. The detection limit and working curves available with test-kits should be confirmed in prior of the use.

When screening contaminated sites for Cr(VI) by test-kit detection, Cr(VI) is extracted from soil into water. The recovery of extracted Cr(VI) would be slightly lower than that when extracting with alkali solution; however, the repeatability of test-kit detection for Cr(VI) is high enough to apply the detection manner to contaminated sites as a screening method. Basically, laboratory analysis provides with precise concentration values but takes longer time and higher costs than those of test-kit detection. Furthermore, under the conditions that survey time and the total costs are restricted, test-kit detection is more useful and practical as it can cover more inspection spots, resulting in precise information on the pollution at investigation sites, compared with a conventional manner applying samples to laboratory analysis that costs and limits the numbers of samples to be analysed due to a lean budget for financial reasons.

The two investigation manners, or the conventional methods and the screening work, have their specific advantages and characters. Then, it is quite natural to choose one of the two options. In short, one gives precise concentration values at selected spots within a limited number but calling for high costs while the other provides concentration levels at spots with no limits in number but showing good cost performance.

This Technical Report describes the procedure to screen soil for Cr(VI) using a method with test-kits developed for water samples.

Soil quality — Detection of water soluble chromium(VI) using a ready-to-use test-kit method

1 Scope

This Technical Report describes the procedure to screen soil samples to detect Cr(VI) using test-kits based on water extraction of Cr(VI) in soil. The test-kit approach in this method is designed to quickly screen soil samples using calibration to indicate the concentration level. A batch test at a liquid to solid ratio of 10 l/kg, designated by ISO/TS 21268-2, has been used in Japan for over 30 years to evaluate the risk of Cr(VI) contamination in soil. The Cr(VI) detection by a ready-to-use test-kit method follows ISO/TS 21268-2.

This screening technique is applicable in laboratories but can also be applied for site screening in the field. The standard system detection covers the range mg/kg to % for Cr(VI).

This method is expected as a screening technique because of the significant colour change given by 1,5-diphenylcarbazine in the existence of Cr(VI). The method may be especially useful in quickly detecting Cr(VI) where a site is assumed to have no Cr(VI) contamination. As with other screening techniques, it is advisable to confirm a certain percentage of both positive and negative test results in another technique, especially when the detected level is near or above a regulatory action limit or when the presence of background or interfering materials is suspected.

Basically, laboratory analysis requires longer time and higher costs than those of test-kit detection. If the same survey time and the costs are allowed with test-kit screening work and conventional investigation methods including laboratory analysis, the former way can work on more inspection spots than the latter. However, test-kit detection should carefully be adopted for site investigation because the recovery of Cr(VI) from soil into water dependent on soil matrices changes the performance of test-kit detection.

2 References

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11074, *Soil quality — Vocabulary*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11074 and the following apply.

3.1

ready-to-use method

analytical method that is ready-made for use, and may be employed in the field

Note 1 to entry: A more familiar name is “field method”.

[SOURCE: ISO 17381:2003, 3.2, modified — definition shortened]

4 Principle

Chromium(VI) in solution reacts with 1,5-diphenylcarbazine to form a red-violet complex of chromium-1,5, diphenylcarbazine resulting in colour changes of the solution due to the formation. Spectrometry and colour charts are applied to determine the complex. The test determines Cr(VI)