



BSI Standards Publication

**Soil quality — Determination
of trace elements in aqua
regia and nitric acid digests
— Graphite furnace atomic
absorption spectrometry
method (GFAAS)**

National foreword

This Published Document is the UK implementation of ISO/TS 17073:2013.

The UK participation in its preparation was entrusted to Technical Committee EH/4, Soil quality.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2013. Published by BSI Standards Limited 2013

ISBN 978 0 580 74393 1

ICS 13.080.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 October 2013.

Amendments issued since publication

Date	Text affected
------	---------------

TECHNICAL SPECIFICATION

ISO/TS 17073

First edition
2013-10-15

Soil quality — Determination of trace elements in aqua regia and nitric acid digests — Graphite furnace atomic absorption spectrometry method (GFAAS)

Qualité du sol — Détermination des éléments en traces solubles dans l'eau régale et l'acide nitrique — Spectrométrie d'absorption atomique avec four graphite



Reference number
ISO/TS 17073:2013(E)

© ISO 2013



COPYRIGHT PROTECTED DOCUMENT

© ISO 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Principle	1
4 Interferences and sources of errors	2
5 Reagents	2
6 Apparatus	4
6.1 Usual laboratory apparatus.....	4
6.2 Atomic absorption spectrometer.....	4
6.3 Automated sample introduction system.....	4
6.4 Graphite tubes.....	4
6.5 Chemical modification.....	4
7 Procedure	5
7.1 Graphite furnace programme.....	5
7.2 Test sample solution.....	5
7.3 Test blank solution.....	6
7.4 Determination.....	6
7.5 Calibration.....	6
8 Calculation	7
9 Expression of results	7
10 Test report	7
Annex A (informative) Examples of instrumental parameter settings	8
Bibliography	9

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical methods and soil characteristics*.

Introduction

ISO/TS 17073 is based upon CEN/TS 16172 *Sludge, treated biowaste and soil — Determination of elements using graphite furnace atomic absorption spectrometry (GF-AAS)*, which was developed by CEN/TC 400, Project Committee — *Horizontal standards in the fields of sludge, biowaste and soil*.

This Technical Specification is applicable and validated for several types of matrices as indicated in [Table 1](#).

Table 1 — Matrices for which this Technical Specification is applicable and validated

Matrix	Materials used for validation
Sludge	Municipal sludge
Biowaste	Compost Fresh compost
Soil	Agricultural soil

Soil quality — Determination of trace elements in aqua regia and nitric acid digests — Graphite furnace atomic absorption spectrometry method (GFAAS)

WARNING — Persons using this Technical Specification should be familiar with usual laboratory practice. This Technical Specification does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this Technical Specification be carried out by suitably trained staff.

1 Scope

This Technical Specification specifies the determination of trace elements in *aqua regia* or nitric acid digests or other extraction procedures of sludge, treated biowaste and soil, using atomic absorption spectrometry with electrothermal atomization in a graphite furnace. The method is applicable for the determination of the following elements:

Arsenic (As), cadmium (Cd), cobalt (Co), lead (Pb), antimony (Sb), thallium (Tl), vanadium (V).

This method may be applied to other elements. The lower working range is approximately 0,01 mg/kg to 0,001 mg/kg, depending on the element to be determined.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 11466, *Soil quality — Extraction of trace elements soluble in aqua regia*

ISO 16729, *Soil quality — Digestion of nitric acid soluble fractions of elements*

ISO 16965, *Soil quality — Determination of trace elements using inductively coupled plasma mass spectrometry (ICP-MS)*

3 Principle

Graphite furnace atomic absorption spectrometry (GFAAS) (also known as Electrothermal Atomic Absorption Spectrometry (ETAAS)), discrete sample aliquots are dispensed into a graphite tube (of which there are several types), which can be heated to over 2 500 °C very rapidly and in a controlled manner. By increasing the temperature stepwise, the processes of drying, thermal decomposition of the matrix and thermal dissociation into free atoms occurs. Atomic absorption spectrometry is based on the ability of free atoms to absorb light. A light source emits light specific for a certain element (or elements). When the light beam passes through the atom cloud in the heated graphite furnace, the light is selectively absorbed by atoms of the chosen element(s). The decrease in light intensity is measured with a detector at a specific wavelength. The concentration of an element in the sample is determined by comparing the absorbance of the sample with the absorbance of calibration solutions. The signal-peak produced is, under optimum conditions, sharp and symmetrical, and of narrow half-width. The peak area is for most elements proportional to the concentration of the element in solution. The measurements are made at the wavelengths given in [Table 2](#).